

X24A SAFETY CHECKLIST			BROOKHAVEN NATIONAL LABORATORY NATIONAL SYNCHROTRON LIGHT SOURCE
Rev: C	Effective: 1/1/05	Page 1 of 1	Number: LS-SCL-0054
Reviewed by: Thomas McDonald		Reviewed by:	Approved by: A. Ackerman

Original signatures on file.

The only official copy of this file is the one on-line in the NSLS Quality Assurance website. Before using a printed copy, verify that it is the most current version by checking the document effective date on the NSLS QA website.

- ☐ 1. Bremsstrahlung shield (BS #1) between GV1 and Mirror Box #1 is banded as per photo.
- ☐ 2. Exclusion Zone # 1 (EZ#1) between monochromator and GV2, defined by polystyrene is identified and secured.
- ☐ 3. The 6" viewport located on the side of monochromator tank is covered by lead glass.
- ☐ 4. Exclusion Zone # 2 (EZ#2) between monochromator and GV4, defined by polystyrene is identified and secured.
- ☐ 5. Bremsstrahlung shield (BS # 2) between the monochromator and GV4 in place and banded as per photo.

Checked by: _____ Date: _____

OPCO: _____ Date: _____